

Nome Prodotto

Compressore Devicare R7E Plus.

Destinazione d'uso

La combinazione di materasso e compressore crea un dispositivo medico per la prevenzione e il sollievo dalle piaghe da decubito in soggetti allettati; sia a domicilio che in ambiente ospedaliero. L'alternarsi dei punti di contatto tra epidermide e materasso effettua un leggero massaggio; si va così a stimolare il flusso sanguineo ed una buona ossigenazione dei tessuti.

**Caratteristiche**

Codice	05.03.R7E-3F - AUTODETECT SYSTEM
AUTODETECT SYSTEM	Sistema digitale di rilevazione automatica del peso e della posizione
Materassi Abbinabili	05.03.K10PLUS • 05.03.K20PLUS
Dimensioni	27 x 22 x 12 cm
Scocca	ABS Ignifugo
Modalità di Erogazione dell'Aria	Statica e Alternata (A-b-A-b)
Modalità 3F	Funzione che mantiene la base ad aria sempre gonfia (statica)
Portata	10 LPM
Campo di Pressione	Da 30 a 150 mm Hg
Rumorosità	≤ 20 Dba
Minuti per Ciclo	10 min, 15 min, 20 min, 25 min
Alimentazione	220/240 V – 50 Hz
Consumo	25 Watt
Lunghezza Cavo Elettrico	4 m, Staccabile
Peso	2.8 Kg
Fusibile	1.5 A

Dotato di due ganci in acciaio per l'ancoraggio al letto e di filtro per l'aria

Allarme visivo e sonoro per bassa pressione e mancanza di elettricità

Avvertenze

Il compressore può essere utilizzato solo con i materassi consigliati dal fabbricante

Verificare il corretto assemblaggio del dispositivo con particolare attenzione al collegamento dei tubi tra compressore e materasso

Manutenzione e Pulizia

Pulire il compressore con un panno umido e detergente neutro

ATTENZIONE: non immergere mai il compressore in alcun tipo di liquido

Conservazione e Smaltimento

Si consiglia l'uso ad un solo utente

Conservare al riparo da fiamme libere e fonti di calore

Smaltire il dispositivo rispettando le normative vigenti in materia di tutela dell'ambiente e raccolta differenziata

Confezione e Nr. Articoli

Scatola 1 Pezzo

Garanzia

Europea 24 mesi per difetti di fabbricazione ad esclusione delle parti consumabili

Riferimenti del Prodotto

Fabbricante Demarta Virginio S.r.l.

Classe di Dispositivo MDR 2017/745 Classe I

Codice ISO 04.33.06.018

Codice CND V080701

Codice RDM 113866/R

Product Name

Devicare R7E Plus Pump.

Use

The combination of mattress and pump creates a medical device for the prevention and relief of pressure ulcers in patients who have to stay in bed for a long time. It is suitable to use both at home or in a hospital. The alternation of the contact points between the skin and the mattress makes a light massage; in this way it is possible to stimulate the blood flow and a good oxygenation of the tissues.



Features	
Code	05.03.R7E-3F - AUTODETECT SYSTEM
AUTODETECT SYSTEM	Digital system of automatic weight and patient position detection
Mattresses Pairing	05.03.K10PLUS • 05.03.K20PLUS
Dimensions	27 x 22 x 12 cm
Case	Fire Retardant ABS
Air Mode	Static and Alternate (A-b-A-b)
3F Mode	Function that keeps the air base always inflated (static)
Air Output	10 LPM
Pressure Range	30 to 150 mm Hg
Noise Level	≤20 Dba
Cycle Time	10 min, 15 min, 20 min, 25 min
Power	220/240 V-50 Hz
Power Consumption	25 Watt
Electric Cable Length	4 m, Detachable
Weight	2.8 Kg
Fuse	1.5 A
Two steel hooks for anchoring to the bed and air filter	
Visual and sound alarm for low pressure and blackout	
Warnings	
The pump can be used only with mattresses recommended by the manufacturer	
Check the correct assembly of the device with particular attention to the connection of the tubes between the pump and the mattress	
Maintenance and Cleaning	
Clean the pump with a damp cloth and neutral detergent	
WARNING: never immerse the compressor in any type of liquid	
Storage and Disposal	
Single-user Product	
Keep away from heat, sparks, open flame or other sources of ignition	
Dispose of the device in compliance with current regulations on environmental protection and separate waste collection	
Package and Nr. Items	
Box	1 Piece
Warranty	
European	24 months for manufacturing defects with the exception of consumable parts
Product References	
Manufacturer	Demarta Virginio S.r.l.
Directive MDR 2017/745	Class I
ISO Code	04.33.06.018
CND Code	V080701
RDM code	113866/R